


Basis Z

Technische Spezifikationen



hydrophob

A-Konstante*	Holladay 1*
119,033 (SRK/T)	1,809 (SF)
Haigis Konstanten*	Holladay 2
a0 = 0,8751	5,57 (ACD)
a1 = 0,2659	Barrett*
a2 = 0,1416	1,90 (LF); N/A (DF)
Hoffer Q*	*Stand IOL Con 03.2026
5,60 (pACD)	
Refraktiver Index	Abbe-Zahl
1,47	58
Linsentyp	Sterilisation
1-teilige Hinterkammerlinse	EO
Material	Empf. Injektionssystem
biokompatibles hydrophobes Acryl (<0,5%)	1stInject 2.0HB, 1stInject 2.4HB
Haptik	Filter
Z-Haptik, 0° Anwinkelung	UV-Filter
Optik	
sphärische Optik, 360° Zellbarriere	

Lieferbereich

Artikelnummer	Sphäre	Schritte
B1AD00	0,0 D bis +9,0 D	1,0 D
B1AD00	+10,0 D bis +30,0 D	0,5 D
S1AD00	+31,0 D bis +35,0 D	1,0 D



Immer die aktuelle Version. QR-Code scannen.
www.1stq.de/basisz1026



Haben Sie noch Fragen?
+49 621 7176330 • info@1stq.de

